



# MEDELSAN

MEDICAL ELECTRONICS

Electronic Eng. Tevfik Soral

**Medikal Cihazlar Üretim – İthalat – Satış ve Teknik Servis Hizmetleri**



## Galvani Faradi Technical Specifications

- G = Galvanic current \* KG = Interrupt Galvanic current Tp= 1 sec Ti = 2 sec
- E1 = Exponancial current Tp = Ti = 1 sec FI = Faramc current Tp = i ms Ti = 2 sec
- S = 50 Hz Sinusoidal current F2 = 100 Hz Faradic current F mod= Modulated F2 with 20 Hz exponancial current
- EElectronic timer is 30 min. And sound warning system
- At the end of the therapy session the output is automatically off
- Supplies a maximum of 100 mA current
- To select easily the requested current with comutator
- The polls can easily be changed by means of the polarity buton

Headquarters: Sakaci Street. 64/20 Kozyatagi / Kadikoy/ Istanbul

Phones: +90 535 736 36 67 / +90 538 642 89 46

Web : [www.medelsan.com](http://www.medelsan.com) E – mail : [info@medelsan.com](mailto:info@medelsan.com)