



MEDELSAN
MEDICAL ELECTRONICS
Electronic Eng. Tefvik Soral

Medikal Cihazlar Üretim – İthalat – Satış ve Teknik Servis Hizmetleri



SHORT WAVE TECHNICAL SPECIFICATIONS

- * Device 220 Vac, 50 Hz also works.
- * Output Frequency 27.12 MHz
- * Output Power 250W
- * TIMER , Digital ,adjustable 0-99 min
- * Continue and Pulse Works
- * Pulse output frequency , 15-26-46-82-150-200 Hz
- * Automatic TUNING
- * There is patient safety circuit
- * At the end of treatment buzzer works
- * There is capacitive- enductive electrods
- * The device is portable
- * Two years guaranteed

Headquarters: Sakaci Str. 64/20 Kozyatagi / Kadikoy / Ist.

Web : www.medelsan.com E – Posta : info@medelsan.com

Phones: +90 535 736 36 67 / +90 538 642 89 46